

Report: CEG.20B170.IB2

Re-issued: 13 March 2020

Page: 1 of 8

Test Report:

EN 1276:2019

Chemical disinfectants and antiseptics – Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas – Test method and requirements (phase 2, step 1)

Identification of the test laboratory:

Abbott Analytical Ltd
Unit 2, Hickmans Road, Birkenhead, CH41 1JH, United Kingdom

Identification of the client:

Centrego Ltd
The Coach House, Newbury, Frome, BA11 3RG, United Kingdom

Identification of the sample:

20B/170

Name of the product:

Toucan ECA Solutions

Batch number/reference and
expiry date (if available):

N/A

Date of delivery:

11 February 2020

Storage conditions:

Room temperature in darkness

Product diluent recommended by
the manufacturer for use:

Not disclosed

Active substance(s) and their
concentrations (s) (optional):

A solution of Hypochlorous acid generated using a single activation of Centrego's Toucan Electrochemical Activation (ECA) system.
Active substances: HOCl and NaOCl.

Appearance of the product:

Clear colourless liquid

Notes:

- 1) The test results in this report relate only to the sample(s) tested.
- 2) This test report may not be reproduced except in full, adapted, altered or used to create a derivative work, without written approval from Abbott Analytical Ltd.
- 3) Re-issue of report CEG.20B170.IB, dated 06 March 2020.
Reason: Addition of active substances and their method of production.

Report: CEG.20B170.IB2

Re-issued: 13 March 2020

Page: 2 of 8

Test method and its validation:

Method: Dilution-neutralisation
Neutraliser: 30.0 g/l Polysorbate 80 + 5.0 g/l Lecithin + 1.0 g/l L-histidine +
1.0 g/l L-cysteine (Neutraliser A)
Neutraliser validation: Validated in accordance with EN 1276:2019 (5.5.2)

Experimental conditions:

Period of analysis: 03 March 2020 to 05 March 2020
Product test concentration(s): Neat
Diluent used for product test solution(s): N/A
Contact time(s): 1 min ± 10 s
Test temperature(s): 20°C ± 1°C
Interfering substance: 0.3 g/l bovine albumin (clean conditions)
Temperature of incubation: 36°C ± 1°C
Identification of the bacterial strain(s) used: *Pseudomonas aeruginosa* (NCIMB 10421)
Escherichia coli (NCTC 10418)
Staphylococcus aureus (NCTC 10788)
Enterococcus hirae (NCIMB 8192)

Deviations: None

Remarks:

- 1) All test conditions are as requested by the client, irrespective of whether these are in accordance with EN 1276:2019 (5.4.2) or EN 1276:2019 (5.5.1.1).
- 2) Products can only be tested at a concentration of 80% or less as some dilution is always produced by adding the test organisms and interfering substance.

Report: CEG.20B170.IB2

Re-issued: 13 March 2020

Page: 3 of 8

Requirements:

The product shall demonstrate at least a 5 decimal log (lg) reduction against every test organism.

Conclusion:

According to EN 1276:2019, this sample of Toucan ECA Solutions possesses bactericidal activity against all of the referenced strains of *Pseudomonas aeruginosa*, *Escherichia coli*, *Staphylococcus aureus* and *Enterococcus hirae*, when tested neat with a contact time of 1 minute at 20°C under clean conditions.

Report re-issued by:

Signed:



Name:

Tony Watson

Position:


General Manager

Date:

13 March 2020

Approved by:

Signed:



Name:

Gareth Bayliss

Position:

Laboratory Manager

Date:

13 March 2020

Report: CEG.20B170.IB2

Re-issued: 13 March 2020

Page: 4 of 8

Results: EN 1276:2019

RST 002 (Issue 4)

Test organism:	<i>Pseudomonas aeruginosa</i>	(NCIMB 10421)
Date of test:	03 March 2020	
Test temperature:	20°C ± 1°C	Incubation temperature: 36°C ± 1°C
Dilution-neutralisation method:	Pour plate	Number of plates: 1 / ml
Neutraliser:	A	Test conditions: Clean conditions

Validation and controls:

Validation suspension (N_{V_0})			Experimental conditions control (A)			Neutraliser or filtration control (B)			Method validation (C) Product conc.: <i>Neat</i>		
Vc1	87	$\bar{x} =$	Vc1	90	$\bar{x} =$	Vc1	91	$\bar{x} =$	Vc1	101	$\bar{x} =$
Vc2	94	90.5	Vc2	87	88.5	Vc2	93	92	Vc2	94	97.5
30 ≤ \bar{x} of N_{V_0} ≤ 160 ?			\bar{x} of A ≥ 0.5 × \bar{x} of N_{V_0} ?			\bar{x} of B ≥ 0.5 × \bar{x} of N_{V_0} ?			\bar{x} of C ≥ 0.5 × \bar{x} of N_{V_0} ?		
<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no		

Test suspension (N and N_0):

N	Vc1	Vc2	\bar{x} wm = 4.15 × 10 ⁸ ;	lg N = 8.62
10 ⁻⁶	>330	>330	$N_0 = N / 10$;	lg N_0 = 7.62
10 ⁻⁷	39	44	7.17 ≤ lg N_0 ≤ 7.70 ?	<input checked="" type="checkbox"/> yes <input type="checkbox"/> no

Test:

Conc. of the product	Contact time	Vc1	Vc2	N_a ($\bar{x} \times 10$)	lg N_a	lg R (lg N_0 - lg N_a)
<i>Neat</i>	1 min	0	0	<140	<2.15	>5.47

Report: CEG.20B170.IB2

Re-issued: 13 March 2020

Page: 5 of 8

Results: EN 1276:2019

RST 002 (Issue 4)

Test organism:	<i>Escherichia coli</i>	(NCTC 10418)
Date of test:	03 March 2020	
Test temperature:	20°C ± 1°C	Incubation temperature: 36°C ± 1°C
Dilution-neutralisation method:	Pour plate	Number of plates: 1 / ml
Neutraliser:	A	Test conditions: Clean conditions

Validation and controls:

Validation suspension (N_{V_0})			Experimental conditions control (A)			Neutraliser or filtration control (B)			Method validation (C) Product conc.: <i>Neat</i>		
Vc1	109	$\bar{x} =$	Vc1	104	$\bar{x} =$	Vc1	120	$\bar{x} =$	Vc1	113	$\bar{x} =$
Vc2	117	113	Vc2	111	107.5	Vc2	110	115	Vc2	107	110
30 ≤ \bar{x} of N_{V_0} ≤ 160 ?			\bar{x} of A ≥ 0.5 × \bar{x} of N_{V_0} ?			\bar{x} of B ≥ 0.5 × \bar{x} of N_{V_0} ?			\bar{x} of C ≥ 0.5 × \bar{x} of N_{V_0} ?		
<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no		

Test suspension (N and N_0):

N	Vc1	Vc2	\bar{x} wm = 2.74×10^8 ; $\lg N = 8.44$	
10^{-6}	276	274	$N_0 = N / 10$; $\lg N_0 = 7.44$	
10^{-7}	27	25	7.17 ≤ $\lg N_0$ ≤ 7.70 ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no	

Test:

Conc. of the product	Contact time	Vc1	Vc2	N_a ($\bar{x} \times 10$)	$\lg N_a$	$\lg R$ ($\lg N_0 - \lg N_a$)
<i>Neat</i>	1 min	0	0	<140	<2.15	>5.29

Report: CEG.20B170.IB2

Re-issued: 13 March 2020

Page: 6 of 8

Results: EN 1276:2019

RST 002 (Issue 4)

Test organism:	<i>Staphylococcus aureus</i>	(NCTC 10788)
Date of test:	03 March 2020	
Test temperature:	20°C ± 1°C	Incubation temperature: 36°C ± 1°C
Dilution-neutralisation method:	Pour plate	Number of plates: 1 / ml
Neutraliser:	A	Test conditions: Clean conditions

Validation and controls:

Validation suspension (N_{V_0})			Experimental conditions control (A)			Neutraliser or filtration control (B)			Method validation (C) Product conc.: <i>Neat</i>		
Vc1	54	$\bar{x} =$	Vc1	61	$\bar{x} =$	Vc1	47	$\bar{x} =$	Vc1	51	$\bar{x} =$
Vc2	50	52	Vc2	50	55.5	Vc2	53	50	Vc2	50	50.5
30 ≤ \bar{x} of N_{V_0} ≤ 160 ?			\bar{x} of A ≥ 0.5 x \bar{x} of N_{V_0} ?			\bar{x} of B ≥ 0.5 x \bar{x} of N_{V_0} ?			\bar{x} of C ≥ 0.5 x \bar{x} of N_{V_0} ?		
<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no		

Test suspension (N and N_0):

N	Vc1	Vc2	\bar{x} wm = 3.60 x 10 ⁸ ;	lg N = 8.56
10 ⁻⁶	>330	>330	$N_0 = N / 10$;	lg N_0 = 7.56
10 ⁻⁷	38	34	7.17 ≤ lg N_0 ≤ 7.70 ?	<input checked="" type="checkbox"/> yes <input type="checkbox"/> no

Test:

Conc. of the product	Contact time	Vc1	Vc2	N_a ($\bar{x} \times 10$)	lg N_a	lg R (lg N_0 - lg N_a)
<i>Neat</i>	1 min	0	0	<140	<2.15	>5.41

Report: CEG.20B170.IB2

Re-issued: 13 March 2020

Page: 7 of 8

Results: EN 1276:2019

RST 002 (Issue 4)

Test organism:	<i>Enterococcus hirae</i>	(NCIMB 8192)
Date of test:	03 March 2020	
Test temperature:	20°C ± 1°C	Incubation temperature: 36°C ± 1°C
Dilution-neutralisation method:	Pour plate	Number of plates: 1 / ml
Neutraliser:	A	Test conditions: Clean conditions

Validation and controls:

Validation suspension (N_{v_0})			Experimental conditions control (A)			Neutraliser or filtration control (B)			Method validation (C) Product conc.: <i>Neat</i>		
Vc1	107	$\bar{x} =$	Vc1	109	$\bar{x} =$	Vc1	104	$\bar{x} =$	Vc1	103	$\bar{x} =$
Vc2	113	110	Vc2	117	113	Vc2	98	101	Vc2	102	102.5
30 ≤ \bar{x} of N_{v_0} ≤ 160 ?			\bar{x} of A ≥ 0.5 × \bar{x} of N_{v_0} ?			\bar{x} of B ≥ 0.5 × \bar{x} of N_{v_0} ?			\bar{x} of C ≥ 0.5 × \bar{x} of N_{v_0} ?		
<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no		

Test suspension (N and N_0):

N	Vc1	Vc2	\bar{x} wm = 4.75 × 10 ⁸ ;	lg N = 8.68
10 ⁻⁶	>330	>330	$N_0 = N / 10$;	lg N_0 = 7.68
10 ⁻⁷	46	49	7.17 ≤ lg N_0 ≤ 7.70 ?	<input checked="" type="checkbox"/> yes <input type="checkbox"/> no

Test:

Conc. of the product	Contact time	Vc1	Vc2	N_a ($\bar{x} \times 10$)	lg N_a	lg R (lg N_0 - lg N_a)
<i>Neat</i>	1 min	0	0	<140	<2.15	>5.53

Explanations:

V_c	count per ml (one plate or more)
\bar{x}	average of V_{c1} and V_{c2} (1 + 2 duplicate)
\bar{x}_{wm}	weighted mean of \bar{x}
N	number of cells per ml in the test suspension
N_0	number of cells in the test mixture at the beginning of the contact time ($N_0 = N / 10$)
N_a	number of survivors per ml in the test mixture at the end of the contact time (before neutralisation or filtration)
R	reduction ($\lg R = \lg N_0 - \lg N_a$)
N_v	number of cells per ml in the validation suspension
N_{v_0}	number of cells in the validation mixtures at the beginning of the contact time ($N_{v_0} = N_v / 10$)
A	number of survivors per ml in the experimental conditions control mixture
B	number of survivors per ml in the neutraliser or filtration control mixture
C	number of survivors per ml in the method validation mixture

All test results have an associated uncertainty of measurement, details of which are available on request.